



NIPDEC

NATIONAL INSURANCE PROPERTY DEVELOPMENT COMPANY LIMITED

June 1, 2010

To All Suppliers

Re: Non Pharmaceutical Tender 2010/2011 – Addendum No. 5 – Request for Clarification

Question 1: **Item numbers 90002442, 90002447, 90002443 and 9000245 - What make and model are the Haemodialysis Machines that the dialyzer will be fitted on?**

Answer: The Haemodialysis Machines in use and for which the Dialysers are required are:

- a. FRISENIUS Brand Machines
- b. BRAUN Brand Machines
- c. BAXTER Brand Machines

The Dialysers provided must be compatible with all three brands.

Question 2: **Please provide the size and thickness required for “Dental Expendables, Pouch Sterilization (Self Seal) 500 pcs”.**

Answer: Dental Expendables – Pouch Sterilization (Self Seal) 500 Box/100
These Pouches are 133mm x 279mm (Size) and 1-3mm (Thickness)

Question 3: **Can you advise on how many wipes are expected to be in 1 carton of “Hard Surface Disinfecting Wipes Antibacterial 2000 Carton 1 Carton”?**

Answer: Hard Surface Disinfecting Wipes Antibacterial – 2000 Cartons

The correct description should read:-

Hard Surface Disinfecting Wipes Antibacterial 2000 Tins x 160 Wipes

The size of the Wipes should be approximately 6” x 6.75”

Please be guided accordingly.

Yours faithfully

Alana Umraw
Company Secretary (Ag.)

HEAD OFFICE

56 - 60 St Vincent St
Port of Spain, Trinidad, WI
TEL: 1 (868) 625-9167, 625-8749 / 50
FAX: 1 (868) 623-0877
E-MAIL: nipdec@tstt.net.tt
WEBSITE: www.nipdec.com

NIPDEC TOBAGO

Unit 7 & 8 Level 3, Block B, NIB Mall,
Scarborough, Tobago, WI
TEL: 1 (868) 639-4118 / 660-7638
FAX: 1 (868) 639-1574

PHARMACEUTICAL DIVISION

C-40 Warehouse, Airways Drive
Chaguaramas, Trinidad, WI
TEL: 1 (868) 634-4406 / 4506 / 4358
FAX: 1 (868) 634-4357
E-MAIL: c40@nipdec.com