



**NIPDEC**

NATIONAL INSURANCE PROPERTY DEVELOPMENT COMPANY LIMITED

---

**May 26, 2010**

**To All Suppliers**

**Re: Non Pharmaceutical Tender 2010/2011 – Addendum No. 2 – Questions**

**1. Question: Item # 90004713 – Indicators, Biological Steam Sterilization- Is this a solution, or what exactly is it?**

**Answer:** This item is used for rapid read out dual biological indicator for steam sterilization. This is used for rapid reliable monitoring of the 270°F (132°C) gravity steam sterilization process. It is used for the detection of the presence of *Geobacillus stearothermophilus*.

**2. Question: Item #90004558-IV Administration: Solution Admin. Set/20G Needle/92' Tube- Most IV Administration sets come without needles, can we therefore supply the IV set and needles as two separate pieces?**

**Answer:** It is preferable to submit a bid for this item as specified.

Please be guided accordingly.

Yours faithfully

Alana Umraw  
Company Secretary (Ag.)